


## Commercial Submission Considerations for Platform Analytical Procedures

#### **Karen Rule**

Director in Analytical Research & Development, Massachusetts USA



#### **Platform Analytical Procedure Development Strategies**





#### The Road Less Travelled...

Approval for a platform analytical procedure requires a large amount of highly subjective justifications vs the traditional approach which is well understood and readily accepted

#### **Submission Content for Traditional Validation**

**Table 1: Typical Performance Characteristics and Related Validation Tests for Measured Quality Attributes** 

Measured Quality Attribute	IDENTITY	IMPURITY (PURITY) Other Quantitative Measurements (1)		ASSAY Content or Potency
Analytical Procedure Performance Characteristics to be Demonstrated (2)		Quantitative Test	Limit Test	Other Quantitative Measurements (1)
Specificity (3) Specificity Test	+	+	+	+
Range Response (Calibration Model)	-	+	-	+
Lower Range Limit	-	$QL^{\dagger}$	DL	-
Accuracy (4) Accuracy Test	-	+	-	+
Precision (4)  Repeatability Test	-	+	-	+
Intermediate Precision Test	-	+(5)	-	+ (5)


ICH Q2(R2) Validation of Analytical Procedures Guidance for Industry; March 2024

#### **Submission Content for Platform Analytical Procedure Validation**

#### Retrospective

- Data
  - Historical or Prior Knowledge
  - Abbreviated Validation for new product
- Series of complicated justifications including:
  - 1. Scientific rationale for why a technology is suitable for a platform
  - 2. Justification for acceptance criteria (to be used for new products) established from historical data
  - 3. Applicability assessment/Risk Assessment for application of platform to a new product
  - 4. Justification for abbreviated validation strategy





# Establishing the Platform Analytical Procedures



#### **Analytical Procedures Targeted for Platform Strategy**

	Quality Attribute	Analytical Procedure		
D	rotein Concentration	UV Spectroscopy		
Г	rotein Concentration	Protein A HPLC (cell culture titer)		
С	harge Profile (Acidic, Main, and Basic Species)	Imaged Capillary Isoelectric Focusing (iCE)		
N	Ionomer and HMMS	Size Exclusion-HPLC (SE-HPLC)		
Н	eavy Chain, Light Chain, and Fragments	Capillary Gel Electrophoresis (CGE, reducing)		
In	ntact IgG and Fragments	Capillary Gel Electrophoresis (CGE, non-reducing)		
Id	lentity	Peptide Map		
Н	ost Cell Protein for DS testing only	HCP ELISA		
N	-Linked Glycan profile for DS testing only	Hydrophilic Interaction Liquid Chromatography		
	HMMS = high molecular mass species, IgG = immunoglobulin G, UV = ultraviolet, HPLC = high performance liquid chromatography; ELISA= enzyme-linked immunosorbent assay			



#### Retrospective Platform Analytical Procedure Product Definition

- All products share a common immunoglobulin structure, similar absorption coefficients and formulations absent of ultraviolet (UV) interference, as well as similar molecular masses (145-149 kDa) and isoelectric points (pl 7.2-9.1), thus enabling the use of common analytical procedures for quality attribute testing
- All products have used the platform analytical procedures without significant changes to the method

Table 1. Example of a Historical Product Library Used to Establish the Analytical Platform

Product	mAb Isoform	Theoretical Mass (kDa)	Main Peak pl	DS Presentation	DP Presentation	Formulation	
			Absorption Coefficient ((mg/mL) <sup>-1</sup> cm <sup>-1</sup> )	Storage Condition	Storage Condition	Formulation	
Historical Prod	lucts <sup>a</sup>	460			1 / "		
mAb-1	laG1	149	7.3	37.0-43.0	9.0-11.0	Platform	
IIIAD-1	lgG1	igai	149	1.39	-20±5 °C	5±3 °C	Platiotili
mAb-2	Bispecific IgG2	148	8.4	38.4-48.0	36.0-44.0	Platform	
IIIAD-2			1.44	-20±5 °C	5±3 °C	Flatioiii	
m 1 h 2	mAb-3 IgG4	149	7.3	142.0-180.0	135.0-165.0	Platform	
MAD-3			1.62	-40±10 °C	5±3 °C	FlatiOfffi	
through mAb-12							

a. mAb-1 through mAb-7 are authorized in the US and EU



#### **UV Spectroscopy Method Established**

#### **Validation Historical Data compiled from 10 mAb's**

Validation Characteristic	Historical Data Summary	Proposed Platform Validation Acceptance Criteria
System Repeatability	RSD: 0.9%	RSD ≤ 3.0%
Method Repeatability	RSD: 1.6%	RSD ≤ 3.0%
Intermediate Precision	RSD: 2.2%	RSD ≤ 5.0%
Reproducibility	RSD: 2.3%	RSD ≤ 5.0%
Accuracy	Accuracy = 102.1%	Accuracy = 100 ± 6%
Specificity	Met Criteria	No response (i.e., slope ≤ 0.1 at 280 nm) is obtained in the formulation buffer while the target sample yields a positive response (i.e., slope > 0.1 at 280 nm)
Linearity	Method is linear over the concentration range (mg/mL) aligned with the molecule's	Response factor (RF) plot shows all points within ± 5% of the average RF
Linearity	specific absorptivity coefficient (mg/ml) <sup>-1</sup> cm <sup>-1</sup>	Data on the linearity plot appears linear by visual inspection
Range	The validated range of the method (mg/mL) is aligned with the molecule's specific absorptivity coefficient (mg/ml) <sup>-1</sup> cm <sup>-1</sup>	Range supports specification acceptance criteria

#### **Critical Operating Parameters Established**

Molecule's Specific Absorption Coefficient, Pathlength

#### **Scientific Rationale**





The ultraviolet absorbance of a protein solution is due to the absorption properties of the aromatic amino acid residues in the protein molecule. According to the Beer-Lambert Law, the concentration of a protein solution is calculated based on the absorbance at a given wavelength, the cuvette cell path length, and the established specific absorption coefficient. The mAb molecules developed share a narrow range of absorption coefficients and formulations absent of UV interference. Therefore, it is feasible to apply the same experimental conditions, pathlength range for variable pathlength or working protein concentration for fixed pathlength, for all the molecules of interest. As a result, the methods can be used from one molecule to another without significant changes to critical method parameters and conditions.



#### **SE-HPLC Method Established**

#### **Validation Historical Data compiled from 10 mAb's**

Validation Characteristic	Historical Data Summary	Proposed Platform Validation Acceptance Criteria
Intermediate Precision	HMMS RSD: 7.25%	HMMS RSD ≤10.0%
Reproducibility	HMMS RSD: 8.9%	HMMS RSD ≤15.0%
Accuracy	HMMS Accuracy = 106.8%	HMMS Accuracy = 100±10%
Specificity	Met Criteria	No response > 0.5% (from the reference material) is obtained in the formulation buffer
Linearity	Method is linear over the range of 0.2-12.5% HMMS	Response factor (RF) plot for HMMS shows all points within ± 20% of the average RF  Data on the linearity plot appears linear by visual inspection
Detection Limit (DL)	0.2 % HMMS	The lowest sample peak area having an s/n ≥ 3
Quantitation Limit (QL)	0.2 % HMMS	The lowest sample peak area having an s/n ≥10
Range	0.2-12.5% HMMS, adjusted based on QL from verification	HMMS = 1% HMMS to approximately 120% of the upper specification limit for %HMMS

#### **Critical Operating Parameters Established**

Column, Mobile Phase, Flow Rate, Sample Preparation, Injection Volume, Column Temperature

#### **Scientific Rationale**





Method uses a X column designed to enable the fractionation of protein species between X,000 and X00,000 Da. Detection is based on UV absorbance at 280 nm with the working protein concentration being the same across products, the absorbances of the mAb molecules of interest are confined in a tight range resulting in consistent method sensitivity. As a result, the method can be used from one molecule to another without significant changes to critical method parameters and conditions. The capability of method to robustly resolve HMMS from monomeric species for mAbs and mAb-like products has been demonstrated through the validation data summarized.




#### **Documentation to Establish Platform Analytical Procedures**

Large investment up front to create this documentation but it is completed once and intended to support all products using the mAb platform analytical procedures

	Document Purpose		Documentation Use in Commercial Submission
	White Paper	High level overarching strategic document to gain strategy alignment across stakeholders, includes the historical product library	Historical product library and overall strategy described in S.4.3 Overview
	Justification/Scientific Rationale	Multi-method justification for platform use, focused from a CQA perspective	Used in summary of method establishment in S.2.6 Evolution of Methods
	Technical Assessment	Contains summary and references for transfers and validations at commercial testing sites	Supportive to S.4.3 and P.5.3
	Platform Validation Workbooks Method Specific	Excel workbooks with all the historical data from validations (one per method)	Submit in Regional to FDA, if asked
	Platform Validation Summary Report – method specific	Summary of scientific rationale, platform validation criteria, results and qualified commercial testing sites (one per method)	Reports Submitted in S.4.3 and/or P.5.3
	Method Development History	Contains summary and references for robustness and critical method operating parameters	Summary showing enhanced development needed in S.2.6 Evolution of Methods




## Application Of Platform to a New Product





#### **Late-Stage Validation Strategy Flow Chart**



<sup>&</sup>quot;Prior product knowledge plays an important role in identifying the appropriate analytical technique. Knowledge of best practices and current state-of-the-art technologies as well as current regulatory expectations contributes to the selection of the most suitable technology for a given purpose. Existing platform analytical procedures (e.g., protein content determination by UV spectroscopy for a protein drug) can be leveraged to evaluate the attributes of a specific product without conducting additional procedure development."

Industry Perspectives on Practical Application of Platform Analytical Procedures. Pharmaceutical Engineering, September/October 2024, 44, Number 5



#### **Applicability Assessment** 3

Analytical procedure has same intended purpose

Analytical procedure used without any significant changes

Analytical procedure evaluations during clinical development align with platform acceptance criteria Comparison to historical library

Assessment of proposed commercial sites

Formulation Buffer

Assessment of proposed commercial specification against prior knowledge for each analytical procedure

#### **New mAb Product**

IgG1, platform formulation + new excipient, 148 kDa, pl 7.4, absorption coefficient 1.55, DS 148-185 mg/mL, DP 139-169 mg/mL

#### Document in S.4.3 and P.5.3

Most complex of all justifications/risk assessment



#### **Applicability Assessment for UV Spectroscopy**

#### Method is Validated and Established as Platform

- √ Validation Summary
- √ Scientific Rationale
- √ Prior knowledge
- √ Critical Method Parameters
- √ Robustness

#### New mAb/-alike Product Applicability Assessment

- ✓ Same intended purpose to measure concentration
- Proposed commercial specification covered by historical validation
- ✓ No critical changes to the procedure

Assessment of validation applicability

Retrospective Analysis of early method evaluation

Extension of Validation – Scientific Rationale

#### **Assess Against Product Library**

- Formulation absent of UV interference
- Specific absorption coefficient determined in the range of previously validated mAbs
- Proposed commercial sites were previously qualified through co-validation or transfer

	ICH Parameters Evaluated	Results	Platform Validation Acceptance Criteria Applied Retrospectively
	Precision (Method)	0.8% 0.9%	Variable Pathlength RSD ≤3.0% Fixed Pathlength RSD ≤5.0%
	Specificity Variable Pathlength Fixed Pathlength	0.00070 ≤0.02 at 280 nm	No response (i.e., slope ≤0.1 at 280 nm) is obtained in the formulation buffer while the target sample yields a positive response (i.e., slope > 0.1 at 280 nm)



#### **Applicability Assessment for SE-HPLC (HMMS)**

### Method is Validated & Established as Platform Validation Summary

- √ Scientific Rationale
- √ Prior knowledge
- √ Critical Method Parameters
- √ Robustness

#### New mAb/-alike Product Applicability Assessment

- ✓ Same intended purpose to measure HMMS
- ✓ Proposed commercial specification covered by historical validation range of 0.2 - 12.5%
- ✓ No critical changes to the procedure

#### Assessment of validation applicability

Retrospective Analysis of early method evaluation

Prepared Sample

Stability

Extension of Validation - Laboratory Verification

Conduct Precision and
Accuracy at the
Commercial Testing Site to
Confirm Expected Results

Predefined acceptance criteria

Protocol

Report

The minimum stability of the

prepared sample for HMMS must

be not less than (NLT) 24 hours

#### **Assess Against Product Library**

- No measurable interference from the formulation
- Mass of product falls within the historical product range validated with this column
- Proposed commercial sites were previously qualified through co-validation or transfer

ICH Parameters Evaluated	Results	Platform Validation Acceptance Criteria Applied Retrospectively
Precision (Method)	0.00%	RSD ≤10.0%
Specificity	No peaks observed between 8-14.5 min	No response > 0.5% (from the reference material) is obtained in the formulation buffer
QL	Confirmed peak response above assay baseline for sample prepared at QL	The lowest sample peak area having an s/n ≥10

Samples are stable

for 24 hr at 2-8C



#### **Summary and Conclusions**

Strategy has been refined over the past year from lessons learned during implementation and from engagement with Health Authorities

#### **Progress & Lessons Learned**

- 1. In Health Authority engagement the overall approach for abbreviated validation was generally acceptable
- Accepted that protein concentration by UV spectroscopy can use extension of validation by scientific rationale
- 3. Historical data from reproducibility or transfers to commercial sites shows the site is qualified to test the platform analytical procedure but agreed any abbreviated validation testing should be conducted at the intended commercial test site
- 4. Better definition of historical mAb products required in submission

#### **Areas for Further Discussion**

- 1. Acceptance Criteria generated from historical data analysis
- 2. Justification and agreement on parameters challenged in abbreviated validation

Need continued discussions to establish a smoother, more predictable path!



#### **Key Challenges**

Documentation of analytical prior knowledge, platform data



Health Authority approval of strategy, globally





Continuous monitoring of platform method performance across sites



Change management of platform method



### Thank You

#### **Acknowledgments:**

#### Core mAb Platform Analytical Procedure Strategy Team

Hetal Patel Julie Armitstead Mike Jankowski Aimee Nichol Melissa Allen Jessea Campbell Scott Sprague Yolli Araujo Tim Graul Karen Rule

#### **Leadership Support**

Jason Starkey Tanya Shang Meg Ruesch

#### **Implementation Team**

Carlos Cabrera Ruiz Matt Thompson Vicky Hayes Dries Vercauteren Kenneth Bruneel Pooja Arora Edel Shannon

